



**KENYATTA UNIVERSITY TEACHING, REFERRAL & RESEARCH HOSPITAL
SUPPLY CHAIN MANAGEMENT**

PO Box 7674- 00100 Nairobi

Cell Phone: 0710642513/0780900519 Email: procurement@kutrrh.go.ke

DATE: 09/11/2022

RE: ADDENDUM NO. 1 TO TENDER NUMBER: KUTRRH /TNR/G/019/ SDME/2022-2023, SUPPLY AND DELIVERY OF MEDICAL EQUIPMENT.

The above is to provide clarification on the following of the above-mentioned tender:

Clarification/Amendment:"

1. The below is additional information for the above tender.

TECHNICAL SPECIFICATIONS PATIENT BEDS (3 CRANK POWERED BED + MATTRESS MACKINTOSH)- Qty 60

Patient Beds (3 Crank Powered bed + Mattress Mackintosh)	The bed must be Electrically operated for adjustment of all positions.	20 marks
	-Four section profiling with perforated steel sheet.	
	a) Positions	
	1) Back rest 0 - 85°	5 marks
	2) Knee break 0 - 30°	5 marks
	3) Footrest 0 - 45 °	5 marks
	4) Trendelenburg/ Reverse Trendelenburg approx. 25°	5 marks
	Safe working load 200 - 230kg or better	10 marks
	Should have urine bag holder	
	Must have Headboard & footboard – Removable with safety device	5 marks
	Must have Fold away side rail	5 marks
	Height adjustment 35cm to 80cm	5 marks
	Overall length -Approx. 190 to 225cm	
	Overall width -Approx. 90 to 100cm	
	Overall, Height -Approx.35 - 80cm	
	Must have wall bumpers	
	Castors –Four swivel antistatic Castors approx. diameter 125mm with two Lockable	5 marks
Must have infusion pole	5 marks	

	CPR release levers / CPR position button for bed to reach "Zero" position quickly.	5 marks
	Power 100 - 240 VAC, Frq, 50/60Hz and must have backup battery (Motor must of high quality, silent while operating & CE certificate	10 marks
	Provide user manual & Technical Manual	
	Warranty: At least one year	5 marks
	Mattress: <ul style="list-style-type: none"> • Four section impermeable Mackintosh/waterproof long stay mattress of 5inch to 8inches thickness 	5 marks
	Pass marks 75%	

All other terms and conditions remain unchanged.

END OF TENDER ADDENDUM

AHMED DAGANE, RN, BA, BSN, MBA-HCM
CHIEF EXECUTIVE OFFICER

JM/pn