



KENYATTA UNIVERSITY TEACHING, REFERRAL & RESEARCH HOSPITAL  
SUPPLY CHAIN MANAGEMENT  
PO Box 7674- 00100 Nairobi  
Cell Phone: 0710642513/0780900519 Email: [procurement@kutrrh.go.ke](mailto:procurement@kutrrh.go.ke)

DATE: 16/11/2023

RE: ADDENDUM NO. 1 :TENDER NUMBER KUTRRH/TNDR/G/012/SDLE/2023-2024, SUPPLY AND DELIVERY OF LABORATORY EQUIPMENT.

The above is to provide clarification on the **following** of the above-mentioned tender:  
Clarification/Amendment; We write to seek clarification regarding the clauses below.

BENCH TOP CENTRIFUGE.

What is the required speed in RPM? &

What is the required RCF?

Amendment;

1. MAX RPM: 4000
2. MAX RCF: 2500xg

Closing date has been changed from 21<sup>st</sup> November 2023 to 24<sup>th</sup> November 2023.

All other terms and conditions remain unchanged.

**AHMED DAGANE, RN, BA, BSN, MBA-HCM**  
**CHIEF EXECUTIVE OFFICER**  
JM/pn



